


CERTIFICATE OF ANALYSIS

General information

Name	GDNA (MDA-MB-436)
Catalog No.	CBP00385
Lot No.	GBD46960820
Format	Genomic DNA
Buffer	Tris-EDTA
Storage Conditions	4°C
Date of Manufacture	Aug.-04-2021
Expiry	36 months from the date of manufacture

Quality Control


Test Items	Standard	Results	Passed
Quantity	5.00 µg/vial*2 vials	5.00 µg/vial*2 vials	√
Concentration	40.00 ng/uL±10%	39.30 ng/uL by Qubit 3.0 detection	√
OD ₂₆₀ /OD ₂₈₀	1.7~2.1	2.06 by Nanodrop detection	√
DNA electrophoresis	No Degradation	 <p> 1kb ladder bp ng/5ul 10000 40 8000 40 6000 40 4000 40 2000 100 1500 40 1000 40 500 40 1% agarose M 1 M: Marker 1: GDNA (MDA-MB-436) </p>	√

Caution:

For research use only.

If you have any questions about the Certificate of Analysis, please contact us.

Certified by:  Date:****

 Date:****