


## CERTIFICATE OF ANALYSIS

### General information

Name	Negative control for DNA
Catalog No.	CBP10000D
Lot No.	CBFA23020306
Format	Genomic DNA
Buffer	Tris-EDTA
Storage Conditions	-20°C
Date of Manufacture	2023/02/03
Expiry	36 months from the date of manufacture


### Quality Control

Test Items	Standard	Results	Passed
Quantity	5.00 µg/vial*1 vial	5.00 µg/vial*1 vial	√
Concentration	40.00 ng/uL±10%	40.00 ng/uL by Qubit 3.0 detection	√
OD <sub>260</sub> /OD <sub>280</sub>	1.7~2.1	1.96 by Nanodrop detection	√
DNA electrophoresis	No Degradation	 <p>1kb ladder</p> <p>M: Marker 1: Negative control for DNA</p>	√

### Caution:

For research use only.

If you have any questions about the Certificate of Analysis, please contact us.

Certified by:  Date:\*\*\*\*

 Date:\*\*\*\*