


CERTIFICATE OF ANALYSIS

Negative control

General information

Name:	Negative control
Catalog No.	CBP10000D
Lot No.	GBD04853827
Format	Genomic DNA
Buffer	Tris-EDTA
Storage Conditions	-20°C
Date of Manufacture	Apr.-29-2022
Expiry	36 months from the date of manufacture


Quality Control

Test Items	Standard	Results	Passed
Quantity	1ug/vial*2vials	1ug/vial*2vials	√
Concentration	Qubit 3.0	30 ng/uL±10%	√
OD ₂₆₀ /OD ₂₈₀	1.7~2.0	1.96	√
DNA electrophoresis	No Degradation	 <p>1kb ladder</p> <p>M: Marker 1: Negative control</p>	√

Caution:

For research use only.

If you have any questions about the Certificate of Analysis, please contact us.

Certified by:  Date:****

 Date:****