


CERTIFICATE OF ANALYSIS

General information

Name	Negative control for gDNA
Catalog No.	CBP10000D
Lot No.	GBD61287210
Format	Genomic DNA
Buffer	Tris-EDTA
Storage Conditions	4°C
Date of Manufacture	Aug.-15-2022
Expiry	36 months from the date of manufacture


Quality Control

Test Items	Standard	Results	Passed
Quantity	1.00 µg/vial*20 vials	1.00 µg/vial*20 vials	√
Concentration	40.00 ng/uL±10%	40.40 ng/uL by Qubit 3.0 detection	√
OD ₂₆₀ /OD ₂₈₀	1.7~2.1	1.96 by Nanodrop detection	√
DNA electrophoresis	No Degradation	 <p>1kb ladder</p> <p>M: Marker 1: Negative control for gDNA</p>	√

Caution:

For research use only.

If you have any questions about the Certificate of Analysis, please contact us.

Certified by:  Date:****

 Date:****