



CERTIFICATE OF ANALYSIS

Product information

Name	TP53 p.R249S Reference Standard
Catalog No.	CBP10157
Lot No.	GBD60128483
Format	Genomic DNA
DNA Change	c.747G>T
AA Change	p.R249S
Mutation Type	Substitution - Missense
Zygoty	Homozygous
%AF	100.00%
Transcript	ENST00000269305
Chr position(GRCh37)	chr17:7577534
Copy Number	2
Ploidy	2.51
Buffer	Tris-EDTA
Storage Conditions	4°C
Date of Manufacture	2022/08/31
Expiry	36 months from the date of manufacture


Quality Control

Test Items	Standard	Results	Passed
Quantity	1.00 µg/vial*1 vial	1.00 µg/vial*1 vial	√
Concentration	40.00 ng/uL±10%	39.60 ng/uL by Qubit 3.0 detection	√
OD ₂₆₀ /OD ₂₈₀	1.7~2.1	1.88 by Nanodrop detection	√
DNA electrophoresis	No Degradation	 <p>1kb ladder</p> <p>M: Marker 1: TP53 p.R249S Reference Standard</p>	√
Genotype	Positive	 <p>By sanger sequencing</p>	√

Caution:

For research use only.

If you have any questions about the Certificate of Analysis, please contact us.

Certified by:  Date:****

 Date:****