


CERTIFICATE OF ANALYSIS

CTNNB1 p.S33C Reference Standard

Product information

Name	CTNNB1 p.S33C Reference Standard
Catalog No.	CBP10250
Lot No.	CGD54065984
Format	Genomic DNA
DNA Change	c.98C>G
AA Change	p.S33C
Mutation Type	Substitution - Missense
Zygoty	Heterozygous
Allelic Frequency	33.33%
Transcript	ENST00000349496
Copy Number	3
Ploidy	3.49
Buffer	Tris-EDTA
Storage Conditions	-20°C
Date of Manufacture	May-10-2020
Expiry	36 months from the date of manufacture


Quality Control

Test Items	Standard	Results	Passed
Quantity	5ug/vial*1vial	5ug/vial*1vial	√
Concentration	Qubit 3.0	40 ng/uL±10%	√
OD ₂₆₀ /OD ₂₈₀	1.7~2.0	1.77	√
DNA electrophoresis	No Degradation	 <p>1kb ladder</p> <p>bp ng/5μL</p> <p>10000 40</p> <p>8000 40</p> <p>6000 40</p> <p>4000 40</p> <p>3000 40</p> <p>2000 100</p> <p>1500 40</p> <p>1000 40</p> <p>500 40</p> <p>1% agarose</p> <p>M</p> <p>1</p> <p>M: Marker</p> <p>1: CTNNB1 p.S33C Reference Standard</p>	√

Caution:

For research use only.

If you have any questions about the Certificate of Analysis, please contact us.

Certified by:  Date:****

 Date:****