



CERTIFICATE OF ANALYSIS

KRAS p.Q61H Reference Standard

Product information

Name	KRAS p.Q61H Reference Standard
Catalog No.	CBP10301
Lot No.	CGD57015516
Format	Genomic DNA
DNA Change	c.183A>T
AA Change	p.Q61H
Mutation Type	Substitution - Missense
Zygoty	Homozygous
Allelic Frequency	100%
Transcript	ENST00000311936
Copy Number	2
Ploidy	2.54
Buffer	Tris-EDTA
Storage Conditions	-20°C
Date of Manufacture	Sep.-23-2019
Expiry	36 months from the date of manufacture


Quality Control

Test Items	Standard	Results	Passed
Quantity	1ug/vial*1vial	1ug/vial*1vial	√
Concentration	Qubit 3.0	40 ng/uL±10%	√
OD ₂₆₀ /OD ₂₈₀	1.7~2.0	1.86	√
DNA electrophoresis	No Degradation	 <p>M: Marker 1: KRAS p.Q61H Reference Standard</p>	√
Genotype	Sanger Sequencing		√
%AF	100%	99.37% by DdPCRdetection	√

Caution:

For research use only.

If you have any questions about the Certificate of Analysis, please contact us.

Certified by:  Date:****

 Date:****