Tel: 4008-750-250

QQ: 4008-750-250 Email: sales@cobioer.com



CERTIFICATE OF ANALYSIS

Product information

Name	NRAS_G12D
Catalog No.	CBP10314
Lot No.	GBD44693131
Format	Genomic DNA
DNA Change	c.35G>A
AA Change	p.G12D
Mutation Type	Substitution - Missense
Zygosity	Heterozygous
%AF	66.67%
Transcript	ENST00000369535
Chr position(GRCh37)	chr1:115258747
Copy Number	3
Ploidy	2.13
Buffer	Tris-EDTA
Storage Conditions	-20°C
Date of Manufacture	Sep13-2022
Expiry	36 months from the date of manufacture

Quality Control

Quality Control			
Test Items	Standard	Results	Passed
Quantity	1.00 μg/vial*1 vial	1.00 μg/vial*1 vial	V
Concentration	40 ng/uL±10%	41 ng/uL	
OD ₂₆₀ /OD ₂₈₀	1.7~2.1	2.06	
DNA electrophoresis	No Degradation	Ikb ladder	√
Genotype	Positive	WWW. GGTGGTGGGGGGGGGGGGGGGGGGGGGGGGGGGGG	√
%AF	66.67%	66.19% by DdPCR detection	





Caution:

For research use only.

If you have any questions about the Certificate of Analysis, please contact us.

Certified by: Date:***

Date:****

Tel: 4008-750-250

QQ: 4008-750-250

Email: sales@cobioer.com