



# CERTIFICATE OF ANALYSIS

## PIK3CA p.H1047L Reference Standard

### Product information

Name	PIK3CA p.H1047L Reference Standard
Catalog No.	CBP10320
Lot No.	GBD58764226
Format	Genomic DNA
DNA Change	c.3140A>T
AA Change	p.H1047L
Mutation Type	Substitution - Missense
Zygoty	Heterozygous
Allelic Frequency	50%
Transcript	NM_006218.1
Copy Number	2
Ploidy	N/A
Buffer	Tris-EDTA
Storage Conditions	-20°C
Date of Manufacture	Nov.-10-2020
Expiry	36 months from the date of manufacture

### Quality Control

Test Items	Standard	Results	Passed
Quantity	1ug/vial*5vials	1ug/vial*5vials	√
Concentration	Qubit 3.0	40 ng/uL±10%	√
OD <sub>260</sub> /OD <sub>280</sub>	1.7~2.0	1.82	√
DNA electrophoresis	No Degradation	 <p>M: Marker I: PIK3CA p.H1047L Reference Standard</p>	√
Genotype	Sanger Sequencing		√
%AF	50%	50.13% by DdPCR detection	√


### Caution:

No.222 Changhong Road, Yuhua  
 District, Nanjing, Jiangsu  
 www.cobioer.com

Tel: 4008-750-250  
 QQ: 4008-750-250  
 Email: sales@cobioer.com

For research use only.

If you have any questions about the Certificate of Analysis, please contact us.

Certified by:  Date:\*\*\*\*

 Date:\*\*\*\*