


CERTIFICATE OF ANALYSIS


PIK3CA p.E545K Reference Standard

Product information

Name	PIK3CA p.E545K Reference Standard
Catalog No.	CBP10322
Lot No.	CGD76946711
Format	Genomic DNA
DNA Change	c.1633G>A
AA Change	p.E545K
Mutation Type	Substitution - Missense
Zygoty	Heterozygous
Allelic Frequency	50%
Transcript	NM_006218.1
Copy Number	2
Ploidy	2.54
Buffer	Tris-EDTA
Storage Conditions	-20°C
Date of Manufacture	May-09-2019
Expiry	36 months from the date of manufacture

Quality Control



Test Items	Standard	Results	Passed
Quantity	5ug/vial*1vial	5ug/vial*1vial	√
Concentration	Qubit 3.0	40 ng/uL	√
OD ₂₆₀ /OD ₂₈₀	1.7~2.0	1.90	√
DNA electrophoresis	No Degradation	 <p> 1kb ladder bp ng/5µl 10000 40 8000 40 6000 40 4000 100 4000 40 3000 40 2000 100 1500 40 1000 40 500 40 1% agarose M 1 M: Marker 1: PIK3CA p.E545K Reference Standard </p>	√



Caution:

For research use only.

If you have any questions about the Certificate of Analysis, please contact us.

Certified by:  Date:****
 Date:****