


CERTIFICATE OF ANALYSIS


PIK3CA p.E545A Reference Standard

Product information

Name	PIK3CA p.E545A Reference Standard
Catalog No.	CBP10323
Lot No.	CGD95483205
Format	Genomic DNA
DNA Change	c.1634A>C
AA Change	p.E545A
Mutation Type	Substitution - Missense
Zygoty	Heterozygous
Allelic Frequency	50%
Transcript	NM_006218.1
Copy Number	2
Ploidy	2.09
Buffer	Tris-EDTA
Storage Conditions	-20°C
Date of Manufacture	Sep.-10-2019
Expiry	36 months from the date of manufacture

Quality Control


Test Items	Standard	Results	Passed
Quantity	1ug/vial*1vial	1ug/vial*1vial	√
Concentration	Qubit 3.0	73 ng/uL	√
OD ₂₆₀ /OD ₂₈₀	1.7~2.0	1.83	√
DNA electrophoresis	No Degradation	 <p>1kb ladder</p> <p>bp ng/5µl</p> <p>10000 40</p> <p>8000 40</p> <p>6000 40</p> <p>4000 40</p> <p>3000 40</p> <p>2000 100</p> <p>1500 40</p> <p>1000 40</p> <p>500 40</p> <p>M</p> <p>1</p> <p>M: Marker</p> <p>1: PIK3CA p.E545A Reference Standard</p>	√

Genotype	Sanger Sequencing		√
----------	-------------------	--	---

Caution:

For research use only.

If you have any questions about the Certificate of Analysis, please contact us.

Certified by:  Date:****

 Date:****