



CERTIFICATE OF ANALYSIS

PIK3CA p.E542K Reference Standard

Product information

Name	PIK3CA p.E542K Reference Standard
Catalog No.	CBP10324
Lot No.	CGD34886856
Format	Genomic DNA
DNA Change	c.1624G>A
AA Change	p.E542K
Mutation Type	Substitution - Missense
Zygoty	Homozygous
Allelic Frequency	100%
Transcript	NM_006218.1
Copy Number	3
Ploidy	2.93
Buffer	Tris-EDTA
Storage Conditions	-20°C
Date of Manufacture	May-09-2019
Expiry	36 months from the date of manufacture


Quality Control

Test Items	Standard	Results	Passed
Quantity	1ug/vial*1vial	1ug/vial*1vial	√
Concentration	Qubit 3.0	100 ng/uL	√
OD ₂₆₀ /OD ₂₈₀	1.7~2.0	1.79	√
DNA electrophoresis	No Degradation	 <p>1kb ladder</p> <p>bp ng/5ul</p> <p>10000 40</p> <p>8000 40</p> <p>6000 40</p> <p>5000 100</p> <p>4000 40</p> <p>3000 40</p> <p>2000 100</p> <p>1500 40</p> <p>1000 40</p> <p>500 40</p> <p>1% agarose</p> <p>M 1</p> <p>M: Marker</p> <p>1: PIK3CA p.E542K Reference Standard</p>	√
Genotype	Sanger Sequencing		√

Caution:

For research use only.

If you have any questions about the Certificate of Analysis, please contact us.

Certified by:  Date:****

 Date:****