


CERTIFICATE OF ANALYSIS


PIK3CA p.E542K Reference Standard

Product information

Name	PIK3CA p.E542K Reference Standard
Catalog No.	CBP10324
Lot No.	GBD88870583
Format	Genomic DNA
DNA Change	c.1624G>A
AA Change	p.E542K
Mutation Type	Substitution - Missense
Zygoty	Homozygous
Allelic Frequency	100%
Transcript	NM_006218.1
Chr position(GRCh37)	chr3:178936082
Copy Number	3
Ploidy	2.93
Buffer	Tris-EDTA
Storage Conditions	4°C
Date of Manufacture	Aug.-02-2022
Expiry	36 months from the date of manufacture

Quality Control


Test Items	Standard	Results	Passed
Quantity	1ug/vial*1vial	1ug/vial*1vial	√
Concentration	Qubit 3.0	40 ng/uL±10%	√
OD ₂₆₀ /OD ₂₈₀	1.7~2.0	1.75	√
DNA electrophoresis	No Degradation	 <p> 1kb ladder bp ng/5ul 10000 40 8000 40 6000 40 5000 100 4000 40 3000 40 2000 100 1500 40 1000 40 500 40 1% agarose M 1 M: Marker 1: PIK3CA p.E542K Reference Standard </p>	√

Genotype	Sanger Sequencing		√
%AF	100%	99.97% by DdPCR detection	√

Caution:

For research use only.

If you have any questions about the Certificate of Analysis, please contact us.

Certified by:  Date:****

 Date:****