


CERTIFICATE OF ANALYSIS


EGFR p.E746_A750delELREA Reference Standard

Product information

Name	EGFR p.E746_A750delELREA Reference Standard
Catalog No.	CBP10334
Lot No.	GBD44271719
Format	Genomic DNA
DNA Change	c.2235_2249del15 GGAATTAAGAGAAGC>-
AA Change	p.E746_A750delELREA
Mutation Type	Deletion - In frame
Zygoty	Heterozygous
Allelic Frequency	88.9%
Transcript	ENST00000275493
Copy Number	6
Ploidy	3.14
Buffer	Tris-EDTA
Storage Conditions	-20°C
Date of Manufacture	Sep.-14-2021
Expiry	36 months from the date of manufacture

Quality Control


Test Items	Standard	Results	Passed
Quantity	1ug/vial*1vial	1ug/vial*1vial	√
Concentration	Qubit 3.0	40 ng/uL±10%	√
OD ₂₆₀ /OD ₂₈₀	1.7~2.0	1.84	√
DNA electrophoresis	No Degradation	 <p> 1kb ladder bp ng/5ul 10000 40 8000 40 6000 40 5000 100 4000 40 3000 40 2000 100 1500 40 1000 40 500 40 1% agarose M 1 M: Marker 1: EGFR p.E746_A750delELREA Reference Standard </p>	√

Genotype	Sanger Sequencing		√
%AF	88.9%	88.7% by DdPCR detection	√

Caution:

For research use only.

If you have any questions about the Certificate of Analysis, please contact us.

Certified by:  Date:****

 Date:****