



CERTIFICATE OF ANALYSIS

EGFR p.S768I Reference Standard

Product information

Name	EGFR p.S768I Reference Standard
Catalog No.	CBP10337
Lot No.	CGD68153629
Format	Genomic DNA
DNA Change	c.2303G>T
AA Change	p.S768I
Mutation Type	Substitution - Missense
Zygoty	Heterozygous
Allelic Frequency	50%
Transcript	ENST00000275493
Copy Number	5
Ploidy	2.27
Buffer	Tris-EDTA
Storage Conditions	-20°C
Date of Manufacture	May-21-2021
Expiry	36 months from the date of manufacture


Quality Control

Test Items	Standard	Results	Passed
Quantity	1ug/vial*1vial	1ug/vial*1vial	√
Concentration	Qubit 3.0	40 ng/uL±10%	√
OD ₂₆₀ /OD ₂₈₀	1.7~2.0	1.82	√
DNA electrophoresis	No Degradation	 <p>1kb ladder</p> <p>bp ng/5µl</p> <p>10000 40</p> <p>8000 40</p> <p>6000 40</p> <p>4000 40</p> <p>3000 40</p> <p>2000 100</p> <p>1500 40</p> <p>1000 40</p> <p>500 40</p> <p>1% agarose</p> <p>M</p> <p>I</p> <p>M: Marker</p> <p>I: EGFR p.S768I Reference Standard</p>	√
Genotype	Sanger Sequencing	 <p>.AGCCTACGTGATGGCCA CGTGGACAACCCCACTGTGCCG</p>	√
%AF	50%	51.4% by DdPCR detection	√

Caution:

For research use only.

If you have any questions about the Certificate of Analysis, please contact us.

Certified by:  Date:****

 Date:****