


CERTIFICATE OF ANALYSIS


AI-Edigen™ EGFR p.C797S Reference Standard Plus

Product information

Name	AI-Edigen™ EGFR p.C797S Reference Standard Plus
Catalog No.	CBP10342
Lot No.	GBD73126206
Format	Genomic DNA
DNA Change	c.2390G>C
AA Change	p.C797S
Mutation Type	Substitution - Missense
Zygoty	Homozygous
Allelic Frequency	100%
Transcript	ENST00000275493
Copy Number	2
Ploidy	2.13
Buffer	Tris-EDTA
Storage Conditions	-20°C
Date of Manufacture	Jul.-27-2020
Expiry	36 months from the date of manufacture

Quality Control


Test Items	Standard	Results	Passed
Quantity	1ug/vial*1vial	1ug/vial*1vial	√
Concentration	Qubit 3.0	40 ng/uL±10%	√
OD ₂₆₀ /OD ₂₈₀	1.7~2.0	1.87	√
DNA electrophoresis	No Degradation	 <p> 1kb ladder bp ng/5µl 10000 40 8000 40 6000 40 4000 100 3000 40 2000 100 1500 40 1000 40 500 40 M: Marker 1: AI-Edigen™ EGFR p.C797S Reference Standard Plus </p>	√

Genotype	Sanger Sequencing		√
%AF	100%	100% by DdPCR detection	√

Caution:

For research use only.

If you have any questions about the Certificate of Analysis, please contact us.

Certified by:  Date: ****

 Date: ****