


CERTIFICATE OF ANALYSIS


EGFR p.747_753LREATSP>S Reference Standard

Product information

Name	EGFR p.747_753LREATSP>S Reference Standard
Catalog No.	CBP10345
Lot No.	CGD48512161
Format	Genomic DNA
DNA Change	c.2240_2257del18 TAAGAGAAGCAACATCTC>-
AA Change	p.L747_P753LREATSP>S
Mutation Type	Deletion - In frame
Zygoty	Heterozygous
Allelic Frequency	25%
Transcript	ENST00000275493
Copy Number	4
Ploidy	2.12
Buffer	Tris-EDTA
Storage Conditions	-20°C
Date of Manufacture	Apr.-09-2021
Expiry	36 months from the date of manufacture

Quality Control


Test Items	Standard	Results	Passed
Quantity	1ug/vial*1vial	1ug/vial*1vial	√
Concentration	Qubit 3.0	40 ng/uL±10%	√
OD ₂₆₀ /OD ₂₈₀	1.7~2.0	1.87	√
DNA electrophoresis	No Degradation	 <p> 1kb ladder bp ng/5ul 10000 40 8000 40 6000 40 4000 40 3000 40 2000 100 1500 40 1000 40 500 40 </p> <p>M: Marker 1: EGFR p.747_753LREATSP>S Reference Standard</p>	√

Genotype	Sanger Sequencing		√
----------	-------------------	------------------------------------------------------------------------------------	---

Caution:

For research use only.

If you have any questions about the Certificate of Analysis, please contact us.

Certified by:  Date:****

 Date:****