



CERTIFICATE OF ANALYSIS

Product information

Name	BRCA1 p.N810T Reference Standard
Catalog No.	CBP10353
Lot No.	CGD90830228
Format	Genomic DNA
DNA Change	c.2429A>C
AA Change	p.N810T
Mutation Type	Missense Mutation
Zygosity	Heterozygous
%AF	50.00%
Transcript	ENST00000357654
Chr position(GRCh37)	chr17:41245119
Copy Number	2
Ploidy	1.93
Buffer	Tris-EDTA
Storage Conditions	4°C
Date of Manufacture	Aug.-30-2021
Expiry	36 months from the date of manufacture


Quality Control

Test Items	Standard	Results	Passed
Quantity	1.00 µg/vial*1 vial	1.00 µg/vial*1 vial	√
Concentration	40.00 ng/uL±10%	41.80 ng/uL by Qubit 3.0 detection	√
OD ₂₆₀ /OD ₂₈₀	1.7~2.1	1.89 by Nanodrop detection	√
DNA electrophoresis	No Degradation	 <p>1kb ladder</p> <p>M: Marker 1: BRCA1 p.N810T Reference Standard</p>	√
Genotype	Positive	 <p>By sanger sequencing</p>	√

Caution:

For research use only.

If you have any questions about the Certificate of Analysis, please contact us.

Certified by:  Date:****

 Date:****