


CERTIFICATE OF ANALYSIS


BRCA2 p.V1862fs*1 Reference Standard

Product information

Name	BRCA2 p.V1862fs*1 Reference Standard
Catalog No.	CBP10385
Lot No.	CGD91685541
Format	Genomic DNA
DNA Change	c.5578delA
AA Change	p.V1862fs*1
Mutation Type	Frame Shift Del
Zygosity	Heterozygous
Allelic Frequency	50%
Transcript	ENST00000380152
Copy Number	2
Ploidy	2.95
Buffer	Tris-EDTA
Storage Conditions	-20°C
Date of Manufacture	Jul.-30-2019
Expiry	36 months from the date of manufacture

Quality Control


Test Items	Standard	Results	Passed
Quantity	1ug/vial*2vials	1ug/vial*2vials	√
Concentration	Qubit 3.0	40 ng/uL±10%	√
OD ₂₆₀ /OD ₂₈₀	1.7~2.0	1.78	√
DNA electrophoresis	No Degradation	 <p style="text-align: center;">M 1</p> <p>M: Marker 1: BRCA2 p.V1862fs*1 Reference Standard</p>	√



Caution:

For research use only.

If you have any questions about the Certificate of Analysis, please contact us.

Certified by:  Date:****

 Date:****