


## CERTIFICATE OF ANALYSIS


### BRCA2 p.N1784fs\*3 Reference Standard

#### Product information

Name	BRCA2 p.N1784fs*3 Reference Standard
Catalog No.	CBP10391
Lot No.	CGD28284951
Format	Genomic DNA
DNA Change	c.5344_5345insA
AA Change	p.N1784fs*3
Mutation Type	Insertion - Frameshift
Zygosity	Heterozygous
Allelic Frequency	50%
Transcript	ENST00000380152
Copy Number	2
Ploidy	2.93
Buffer	Tris-EDTA
Storage Conditions	-20°C
Date of Manufacture	Mar.-20-2019
Expiry	36 months from the date of manufacture

#### Quality Control



Test Items	Standard	Results	Passed
Quantity	1ug/vial*2vials	1ug/vial*2vials	√
Concentration	Qubit 3.0	40 ng/uL±10%	√
OD <sub>260</sub> /OD <sub>280</sub>	1.7~2.0	1.90	√
DNA electrophoresis	No Degradation	 <p>1kb ladder</p> <p>bp ng/5ul</p> <p>10000 40</p> <p>8000 40</p> <p>6000 40</p> <p>4000 40</p> <p>3000 40</p> <p>2000 100</p> <p>1500 40</p> <p>1000 40</p> <p>500 40</p> <p>1% agarose</p> <p>M 1</p> <p>M: Marker</p> <p>1: BRCA2 p.N1784fs*3 Reference Standard</p>	√

Genotype	Sanger Sequencing	
----------	-------------------	--

**Caution:**

For research use only.

If you have any questions about the Certificate of Analysis, please contact us.

Certified by:  Date:\*\*\*\*  
 Date:\*\*\*\*