



CERTIFICATE OF ANALYSIS

Product information

Name	BRCA2.p.N1287fs*6 Reference Standard
Catalog No.	CBP10393
Lot No.	GBD92065720
Format	Genomic DNA
DNA Change	c.3854delA
AA Change	p.N1287fs*6
Mutation Type	Deletion - Frameshift
Zygosity	Heterozygous
%AF	50.00%
Transcript	ENST00000380152
Chr position(GRCh37)	chr13:32912352
Copy Number	2
Ploidy	1.97
Buffer	Tris-EDTA
Storage Conditions	4°C
Date of Manufacture	Jun.-28-2022
Expiry	36 months from the date of manufacture

Quality Control


Test Items	Standard	Results	Passed
Quantity	1.00 µg/vial*1 vial	1.00 µg/vial*1 vial	√
Concentration	40.00 ng/uL±10%	37.90 ng/uL by Qubit 3.0 detection	√
OD ₂₆₀ /OD ₂₈₀	1.7~2.1	1.81 by Nanodrop detection	√
DNA electrophoresis	No Degradation	 <p>M: Marker 1: BRCA2.p.N1287fs*6 Reference Standard</p>	√
Genotype	Positive	 <p>By sanger sequencing</p>	√

%AF	50.00%	49.70% by DdPCR detection	√
-----	--------	---------------------------	---

Caution:

For research use only.

If you have any questions about the Certificate of Analysis, please contact us.

Certified by:  Date:****

 Date:****