



# CERTIFICATE OF ANALYSIS EGFR p.L861Q Reference Standard

## Product information

Name	EGFR p.L861Q Reference Standard
Catalog No.	CBP10401
Lot No.	CGD31559074
Format	Genomic DNA
DNA Change	c.2582T>A
AA Change	p.L861Q
Mutation Type	Substitution - Missense
Zygosity	Homozygous
Allelic Frequency	100%
Transcript	ENST00000275493
Copy Number	2
Ploidy	2.13
Buffer	Tris-EDTA
Storage Conditions	-20°C
Date of Manufacture	May-14-2021
Expiry	36 months from the date of manufacture


## Quality Control

Test Items	Standard	Results	Passed
Quantity	1ug/vial*1vial	1ug/vial*1vial	√
Concentration	Qubit 3.0	40 ng/uL±10%	√
OD <sub>260</sub> /OD <sub>280</sub>	1.7~2.0	1.93	√
DNA electrophoresis	No Degradation	 <p>1kb ladder</p> <p>bp ng/5ul</p> <p>10000 40</p> <p>8000 40</p> <p>6000 40</p> <p>5000 100</p> <p>4000 40</p> <p>3000 40</p> <p>2000 100</p> <p>1500 40</p> <p>1000 40</p> <p>500 40</p> <p>1% agarose</p> <p>M</p> <p>1</p> <p>M: Marker</p> <p>1: EGFR p.L861Q Reference Standard</p>	√
Genotype	Sanger Sequencing		√
%AF	100.00%	99.92% by DdPCR detection	√

**Caution:**

For research use only.

If you have any questions about the Certificate of Analysis, please contact us.

Certified by:  Date:\*\*\*\*

 Date:\*\*\*\*