



CERTIFICATE OF ANALYSIS AI-Edigene® EGFR p.S768I Reference Standard Plus

Product information

Name	AI-Edigene® EGFR p.S768I Reference Standard Plus
Catalog No.	CBP10405
Lot No.	GBD28553855
Format	Genomic DNA
DNA Change	c.2303G>T
AA Change	p.S768I
Mutation Type	Substitution - Missense
Zygosity	Homozygous
Allelic Frequency	100%
Transcript	ENST00000275493
Copy Number	2
Ploidy	2.13
Buffer	Tris-EDTA
Storage Conditions	-20°C
Date of Manufacture	Aug.-05-2020
Expiry	36 months from the date of manufacture



Quality Control

Test Items	Standard	Results	Passed
Quantity	1ug/vial*1vial	1ug/vial*1vial	√
Concentration	Qubit 3.0	40 ng/uL±10%	√
OD ₂₆₀ /OD ₂₈₀	1.7~2.0	1.85	√
DNA electrophoresis	No Degradation	 <p>1kb ladder</p> <p>bp ng/5ul</p> <p>10000 40</p> <p>8000 40</p> <p>6000 40</p> <p>5000 100</p> <p>4000 40</p> <p>3000 40</p> <p>2000 100</p> <p>1500 40</p> <p>1000 40</p> <p>500 40</p> <p>1% agarose</p> <p>M</p> <p>1</p> <p>M: Marker</p> <p>1: AI-Edigene® EGFR p.S768I Reference Standard Plus</p>	√
Genotype	Sanger Sequencing	 <p>G A A G C C T A C G T G A T G G C C A T C G T G G A C A A C C C C C</p>	√
Allelic Frequency	100.00%	99.97% by DdPCR detection	√

Caution:

For research use only.

If you have any questions about the Certificate of Analysis, please contact us.

Certified by:  Date:****
 Date:****