



CERTIFICATE OF ANALYSIS AI-Edigene® BRAF p.V600K Reference Standard Plus

Product information

Name	AI-Edigene® BRAF p.V600K Reference Standard Plus
Catalog No.	CBP10413
Lot No.	CGD22020024
Format	Genomic DNA
DNA Change	c.1798_1799GT>AA
AA Change	p.V600K
Mutation Type	Substitution - Missense
Zygosity	Homozygous
Allelic Frequency	100%
Transcript	ENST00000288602
Chr position(GRCh37)	chr7:140453136_140453137
Copy Number	2
Ploidy	1.98
Buffer	Tris-EDTA
Storage Conditions	4°C
Date of Manufacture	Feb.-10-2022
Expiry	36 months from the date of manufacture

Quality Control


Test Items	Standard	Results	Passed
Quantity	1ug/vial*10vials	1ug/vial*10vials	√
Concentration	Qubit 3.0	25 ng/uL±10%	√
OD ₂₆₀ /OD ₂₈₀	1.7~2.0	1.92	√
DNA electrophoresis	No Degradation	 <p> 1kb ladder bp ng/5ul 10000 40 8000 40 6000 40 5000 100 4000 40 3000 40 2000 100 1500 40 1000 40 500 40 1% agarose M 1 M: Marker 1: AI-Edigene® BRAF p.V600K Reference Standard Plus </p>	√

Genotype	Sanger Sequencing		√
%AF	100%	99.97% by DdPCR detection	√

Caution:

For research use only.

If you have any questions about the Certificate of Analysis, please contact us.

Certified by:  Date:****

 Date:****