



CERTIFICATE OF ANALYSIS

Product information

Name	FLT3-ITD p.D600_L601ins10 Reference Standard
Catalog No.	CBP10421
Lot No.	CBFA23033002X
Format	Genomic DNA
DNA Change	c.1801_1802ins30(ACGTTGATTCAGAGAATATGAATATGATC)
AA Change	p.D600_L601ins10
Mutation Type	Insertion - In frame
Zygoty	Homozygous
%AF	100.00%
Transcript	NM_004119.2
Copy Number	2
Ploidy	2.01
Buffer	Tris-EDTA
Storage Conditions	4°C
Date of Manufacture	2023/03/30
Expiry	36 months from the date of manufacture


Quality Control

Test Items	Standard	Results	Passed
Quantity	1.00 µg/vial*1 vial	1.00 µg/vial*1 vial	√
Concentration	40.00 ng/uL±10%	42.60 ng/uL by Qubit 3.0 detection	√
OD ₂₆₀ /OD ₂₈₀	1.7~2.1	1.90 by Nanodrop detection	√
DNA electrophoresis	No Degradation	 <p>1kb ladder</p> <p>M: Marker 1: FLT3-ITD p.D600_L601ins10 Reference Standard</p>	√
Genotype	Positive	 <p>By sanger sequencing</p>	√
%AF	100.00%	100.00% by DdPCR detection	√

Caution:

For research use only.

If you have any questions about the Certificate of Analysis, please contact us.

Certified by:  Date:****

 Date:****