



CERTIFICATE OF ANALYSIS

BRAF p.V600D Reference Standard Plus

Product information

Name	BRAF p.V600D Reference Standard Plus
Catalog No.	CBP10423
Lot No.	CGD27851907
Format	Genomic DNA
DNA Change	c.1799_1800TG>AT
AA Change	p.V600D
Mutation Type	Substitution - Missense
Zygoty	Homozygous
Allelic Frequency	100%
Transcript	ENST00000288602
Copy Number	2
Ploidy	1.98
Buffer	Tris-EDTA
Storage Conditions	-20°C
Date of Manufacture	Apr.-23-2021
Expiry	36 months from the date of manufacture


Quality Control

Test Items	Standard	Results	Passed
Quantity	1ug/vial*1vial	1ug/vial*1vial	√
Concentration	Qubit 3.0	40 ng/uL±10%	√
OD ₂₆₀ /OD ₂₈₀	1.7~2.0	1.80	√
DNA electrophoresis	No Degradation	 <p>1kb ladder</p> <p>M: Marker 1: BRAF p.V600D Reference Standard Plus</p>	√
Genotype	Sanger Sequencing		√
%AF	100%	100% by DdPCR detection	√

Caution:

For research use only.

If you have any questions about the Certificate of Analysis, please contact us.

Certified by:  Date:****

 Date:****