


CERTIFICATE OF ANALYSIS


AI-Edigene® BRAF p.V600E Reference Standard Plus

Product information

Name	AI-Edigene® BRAF p.V600E Reference Standard Plus
Catalog No.	CBP10424
Lot No.	CGD19671149
Format	Genomic DNA
DNA Change	c.1799T>A
AA Change	p.V600E
Mutation Type	Substitution - Missense
Zygosity	Heterozygous
Allelic Frequency	50%
Transcript	ENST00000288602
Copy Number	2
Ploidy	1.98
Buffer	Tris-EDTA
Storage Conditions	-20°C
Date of Manufacture	Oct.-08-2019
Expiry	36 months from the date of manufacture

Quality Control


Test Items	Standard	Results	Passed
Quantity	1ug/vial*1vial	1ug/vial*1vial	√
Concentration	Qubit 3.0	40 ng/uL±10%	√
OD ₂₆₀ /OD ₂₈₀	1.7~2.0	1.82	√
DNA electrophoresis	No Degradation	 <p>1kb ladder</p> <p>M: Marker 1: AI-Edigene® BRAF p.V600E Reference Standard Plus</p>	√

Genotype	Sanger Sequencing		√
Allelic Frequency	50.00%	50.04% by DdPCR detection	√

Caution:

For research use only.

If you have any questions about the Certificate of Analysis, please contact us.

Certified by:  Date:****

 Date:****