



CERTIFICATE OF ANALYSIS

Product information

Name	TP53 p.165_166insYKQ Reference Standard
Catalog No.	CBP10453
Lot No.	GBD24709386
Format	Genomic DNA
DNA Change	c.496-497insACAAGCAGT
AA Change	p.165_166insYKQ
Mutation Type	In Frame Ins
Zygosity	Unkown
%AF	Unkown
Transcript	ENST00000269305.4
Chr position(GRCh37)	chr17:7578433_7578434
Copy Number	3
Ploidy	2.65
Buffer	Tris-EDTA
Storage Conditions	4°C
Date of Manufacture	Jun.-09-2022
Expiry	36 months from the date of manufacture


Quality Control

Test Items	Standard	Results	Passed
Quantity	1.00 µg/vial*1 vial	1.00 µg/vial*1 vial	√
Concentration	40.00 ng/uL±10%	40.00 ng/uL by Qubit 3.0 detection	√
OD ₂₆₀ /OD ₂₈₀	1.7~2.1	1.86 by Nanodrop detection	√
DNA electrophoresis	No Degradation	 <p>1kb ladder</p> <p>M: Marker 1: TP53 p.165_166insYKQ Reference Standard</p>	√
Genotype	Positive	 <p>By sanger sequencing</p>	√

Caution:

For research use only.

If you have any questions about the Certificate of Analysis, please contact us.

Certified by:  Date:****

 Date:****