



CERTIFICATE OF ANALYSIS

Product information

Name	BRCA2 p.F1427fs Reference Standard
Catalog No.	CBP10454
Lot No.	CBFA23090916X
Format	Genomic DNA
DNA Change	c.4277_4278insT
AA Change	p.F1427fs
Mutation Type	Frame Shift Ins
Zygoty	Heterozygous
%AF	50.00%
Transcript	ENST00000380152.3
Chr position(GRCh37)	chr13:32912770_32912771
Copy Number	2
Ploidy	2.02
Buffer	Tris-EDTA
Storage Conditions	4°C
Date of Manufacture	2023/09/09
Expiry	36 months from the date of manufacture


Quality Control

Test Items	Standard	Results	Passed
Quantity	1.00 µg/vial*2 vials	1.00 µg/vial*2 vials	√
Concentration	40.00 ng/uL±10%	43.20 ng/uL by Qubit 3.0 detection	√
OD ₂₆₀ /OD ₂₈₀	1.7~2.1	1.89 by Nanodrop detection	√
DNA electrophoresis	No Degradation	 <p>1kb ladder</p> <p>M: Marker 1: BRCA2 p.F1427fs Reference Standard</p>	√
Genotype	Positive	 <p>By sanger sequencing</p>	√

Caution:

For research use only.

If you have any questions about the Certificate of Analysis, please contact us.

Certified by:  Date:****

 Date:****