



# CERTIFICATE OF ANALYSIS

## BRCA2 p.F1427fs Reference Standard

### Product information

Name	BRCA2 p.F1427fs Reference Standard
Catalog No.	CBP10454
Lot No.	CGD22390086
Format	Genomic DNA
DNA Change	c.4277_4278insT
AA Change	p.F1427fs
Mutation Type	Frame Shift Ins
Zygoty	Heterozygous
Allelic Frequency	50%
Transcript	ENST00000380152.3
Copy Number	2
Ploidy	2.02
Buffer	Tris-EDTA
Storage Conditions	-20°C
Date of Manufacture	Oct.-23-2019
Expiry	36 months from the date of manufacture


### Quality Control

Test Items	Standard	Results	Passed
Quantity	1ug/vial*2vials	1ug/vial*2vials	√
Concentration	Qubit 3.0	40 ng/uL±10%	√
OD <sub>260</sub> /OD <sub>280</sub>	1.7~2.0	1.89	√
DNA electrophoresis	No Degradation	 <p>M: Marker 1: BRCA2 p.F1427fs Reference Standard</p>	√
Genotype	Sanger Sequencing		√

**Caution:**

For research use only.

If you have any questions about the Certificate of Analysis, please contact us.

Certified by:  Date:\*\*\*\*

 Date:\*\*\*\*