Tel: 4008-750-250

QQ: 4008-750-250 Email: sales@cobioer.com



CERTIFICATE OF ANALYSIS

Product information

AI-Edigene® ERBB2 p.V777_G778insCG Reference Standard Plus			
CBP10474			
CGD28868019			
Genomic DNA			
c.2331_2332insTGTGGG			
p.V777_G778insCG			
In_Frame_Ins			
Homozygous			
100.00%			
NM_004448.3			
chr17:37881002_37881003			
2			
2.13			
Tris-EDTA			
4°C			
Nov21-2020			
36 months from the date of manufacture			

Quality Control

Quanty Control			
Test Items	Standard	Results	Passed
Quantity	1.00 μg/vial*1 vial	1.00 μg/vial*1 vial	
Concentration	40.00 ng/uL±10%	39.30 ng/uL by Qubit 3.0 detection	
OD_{260}/OD_{280}	1.7~2.1	1.74 by Nanodrop detection	
DNA electrophoresis	No Degradation	1kb ladder	√
Genotype	Positive	By sanger sequencing	V





Caution:

For research use only.

If you have any questions about the Certificate of Analysis, please contact us.

Certified by: Date:***

Date:***

Tel: 4008-750-250

QQ: 4008-750-250

Email: sales@cobioer.com