



CERTIFICATE OF ANALYSIS

Product information

Name	AI-Edigene™ KRAS p.G12V Reference Standard Plus
Catalog No.	CBP10477
Lot No.	CGFA23020804
Format	Genomic DNA
DNA Change	c.35G>T
AA Change	p.G12V
Mutation Type	Substitution - Missense
Zygoty	Homozygous
%AF	100.00%
Transcript	ENST00000311936
Chr position(GRCh37)	chr12:25398284
Copy Number	2
Ploidy	2.13
Buffer	Tris-EDTA
Storage Conditions	4°C
Date of Manufacture	Feb.-08-2023
Expiry	36 months from the date of manufacture

Quality Control


Test Items	Standard	Results	Passed
Quantity	1.00 µg/vial*2 vials	1.00 µg/vial*2 vials	√
Concentration	40.00 ng/uL±10%	40.00 ng/uL by Qubit 3.0 detection	√
OD ₂₆₀ /OD ₂₈₀	1.7~2.1	1.77 by Nanodrop detection	√
DNA electrophoresis	No Degradation	 <p>1kb ladder</p> <p>M: Marker 1: AI-Edigene™ KRAS p.G12V Reference Standard Plus</p>	√
Genotype	Positive	 <p>By sanger sequencing</p>	√

%AF	100.00%	99.99% by DdPCR detection	√
-----	---------	---------------------------	---

Caution:

For research use only.

If you have any questions about the Certificate of Analysis, please contact us.

Certified by:  Date:****

 Date:****