


CERTIFICATE OF ANALYSIS


AI-Edigene® KRAS p.G12C Reference Standard Plus

Product information

Name	AI-Edigene® KRAS p.G12C Reference Standard Plus
Catalog No.	CBP10478
Lot No.	CGD09013853
Format	Genomic DNA
DNA Change	c.34G>T
AA Change	p.G12C
Mutation Type	Substitution - Missense
Zygosity	Homozygous
Allelic Frequency	100%
Transcript	ENST00000311936
Copy Number	2
Ploidy	2.13
Buffer	Tris-EDTA
Storage Conditions	-20°C
Date of Manufacture	Jan.-06-2020
Expiry	36 months from the date of manufacture

Quality Control



Test Items	Standard	Results	Passed
Quantity	1ug/vial*5vials	1ug/vial*5vials	√
Concentration	Qubit 3.0	40 ng/uL±10%	√
OD ₂₆₀ /OD ₂₈₀	1.7~2.0	1.87	√
DNA electrophoresis	No Degradation	 <p>1kb ladder</p> <p>bp ng5ul</p> <p>10000 40</p> <p>8000 40</p> <p>6000 40</p> <p>4000 40</p> <p>3000 40</p> <p>2000 100</p> <p>1500 40</p> <p>1000 40</p> <p>500 40</p> <p>M</p> <p>1</p> <p>M: Marker</p> <p>1: AI-Edigene® KRAS p.G12C Reference Standard Plus</p>	√

Genotype	Sanger Sequencing		√
----------	-------------------	--	---

Caution:

For research use only.

If you have any questions about the Certificate of Analysis, please contact us.

Certified by:  Date:****
 Date:****