


CERTIFICATE OF ANALYSIS


AI-Edigene® KRAS p.G13D Reference Standard Plus

Product information

Name	AI-Edigene® KRAS p.G13D Reference Standard Plus
Catalog No.	CBP10481
Lot No.	GBD29524632
Format	Genomic DNA
DNA Change	c.38G>A
AA Change	p.G13D
Mutation Type	Substitution - Missense
Zygoty	Homozygous
Allelic Frequency	100%
Transcript	ENST00000311936
Chr position(GRCh37)	chr12:25398281
Copy Number	2
Ploidy	2.13
Buffer	Tris-EDTA
Storage Conditions	4°C
Date of Manufacture	Sep.-13-2022
Expiry	36 months from the date of manufacture

Quality Control


Test Items	Standard	Results	Passed
Quantity	1ug/vial*8vials	1ug/vial*8vials	√
Concentration	Qubit 3.0	40 ng/uL±10%	√
OD ₂₆₀ /OD ₂₈₀	1.7~2.0	1.88	√
DNA electrophoresis	No Degradation	 <p> 1kb ladder bp ng/5ul 10000 40 8000 40 6000 40 4000 100 4000 40 3000 40 2000 100 1500 40 1000 40 500 40 M: Marker 1: AI-Edigene® KRAS p.G13D Reference Standard Plus </p>	√

Genotype	Sanger Sequencing		√
%AF	100%	100% by DdPCR detection	√

Caution:

For research use only.

If you have any questions about the Certificate of Analysis, please contact us.

Certified by:  Date:****

 Date:****