


CERTIFICATE OF ANALYSIS


AI-Edigene® EGFR p.G719C Reference Standard Plus

Product information

Name	AI-Edigene® EGFR p.G719C Reference Standard Plus
Catalog No.	CBP10500
Lot No.	CGD61156602
Format	Genomic DNA
DNA Change	c.2155G>T
AA Change	p.G719C
Mutation Type	Substitution - Missense
Zygoty	Homozygous
Allelic Frequency	100%
Transcript	ENST00000275493
Copy Number	2
Ploidy	2.13
Buffer	Tris-EDTA
Storage Conditions	-20°C
Date of Manufacture	Mar.-31-2020
Expiry	36 months from the date of manufacture

Quality Control


Test Items	Standard	Results	Passed
Quantity	1ug/vial*5vials	1ug/vial*5vials	√
Concentration	Qubit 3.0	40 ng/uL±10%	√
OD ₂₆₀ /OD ₂₈₀	1.7~2.0	1.88	√
DNA electrophoresis	No Degradation	 <p>1kb ladder</p> <p>M: Marker 1: AI-Edigene® EGFR p.G719C Reference Standard Plus</p>	√

Genotype	Sanger Sequencing		√
Allelic Frequency	100.00%	100.00% by DdPCR detection	√

Caution:

For research use only.

If you have any questions about the Certificate of Analysis, please contact us.

Certified by:  Date:****

 Date:****