


# CERTIFICATE OF ANALYSIS BRCA2 2134bp-Del Reference Standard

## Product information

Name	BRCA2 2134bp-Del Reference Standard
Catalog No.	CBP10505
Lot No.	CGD70482075
Format	Genomic DNA
DNA Change	13:32913403-32915536del2134bp
Transcript	ENST00000380152.3
Copy Number	2
Ploidy	3.52
Buffer	Tris-EDTA
Storage Conditions	-20°C
Date of Manufacture	Oct.-09-2019
Expiry	36 months from the date of manufacture


## Quality Control

Test Items	Standard	Results	Passed
Quantity	1ug/vial*2vials	1ug/vial*2vials	√
Concentration	Qubit 3.0	40 ng/uL	√
OD <sub>260</sub> /OD <sub>280</sub>	1.7~2.0	1.95	√
DNA electrophoresis	No Degradation	 <p>1kb ladder</p> <p>bp ng/μl</p> <p>10000 40</p> <p>8000 40</p> <p>6000 40</p> <p>5000 100</p> <p>4000 40</p> <p>3000 40</p> <p>2000 100</p> <p>1500 40</p> <p>1000 40</p> <p>500 40</p> <p>M: Marker</p> <p>1: BRCA2 2134bp-Del Reference Standard</p>	√

### Caution:

For research use only.

If you have any questions about the Certificate of Analysis, please contact us.

Certified by:  Date:\*\*\*\*

 Date:\*\*\*\*