



CERTIFICATE OF ANALYSIS

Product information

Name	AI-Edigene® c-KIT p.D816V Reference Standard Plus
Catalog No.	CBP10507
Lot No.	GBD17225422
Format	Genomic DNA
DNA Change	c.2447A>T
AA Change	p.D816V
Mutation Type	Substitution - Missense
Zygoty	Homozygous
%AF	100.00%
Transcript	NM_000222.2
Copy Number	2
Ploidy	1.99
Buffer	Tris-EDTA
Storage Conditions	4°C
Date of Manufacture	Jul.-28-2022
Expiry	36 months from the date of manufacture


Quality Control

Test Items	Standard	Results	Passed
Quantity	1.00 µg/vial*1 vial	1.00 µg/vial*1 vial	√
Concentration	35.00 ng/uL±10%	35.00 ng/uL by Qubit 3.0 detection	√
OD ₂₆₀ /OD ₂₈₀	1.7~2.1	1.81 by Nanodrop detection	√
DNA electrophoresis	No Degradation	 <p>1kb ladder</p> <p>M: Marker 1: AI-Edigene® c-KIT p.D816V Reference Standard Plus</p>	√
Genotype	Positive	 <p>By sanger sequencing</p>	√
%AF	100.00%	99.93% by DdPCR detection	√

Caution:

For research use only.

If you have any questions about the Certificate of Analysis, please contact us.

Certified by:  Date: ****

 Date: ****