


CERTIFICATE OF ANALYSIS


AI-Edigene IDH2 p.R172K eference Standard Plus

Product information

Name	AI-Edigene IDH2 p.R172K eference Standard Plus
Catalog No.	CBP10514
Lot No.	GBD98058724
Format	Genomic DNA
DNA Change	c.515G>A
AA Change	p.R172K
Mutation Type	Substitution - Missense
Zygosity	Homozygous
Allelic Frequency	100%
Transcript	NM_002168.4
Copy Number	2
Ploidy	1.99
Buffer	Tris-EDTA
Storage Conditions	-20°C
Date of Manufacture	Nov.-10-2020
Expiry	36 months from the date of manufacture

Quality Control


Test Items	Standard	Results	Passed
Quantity	1ug/vial*1vial	1ug/vial*1vial	√
Concentration	Qubit 3.0	40 ng/uL±10%	√
OD ₂₆₀ /OD ₂₈₀	1.7~2.0	1.90	√
DNA electrophoresis	No Degradation	 <p>1kb ladder</p> <p>M: Marker 1: AI-Edigene IDH2 p.R172K eference Standard Plus</p>	√

Genotype	Sanger Sequencing		√
%AF	100%	99.92% by DdPCR detection	√

Caution:

For research use only.

If you have any questions about the Certificate of Analysis, please contact us.

Certified by:  Date:****

 Date:****