



CERTIFICATE OF ANALYSIS

Product information

Name	AI-Edigene® PIK3CA p.E545K Reference Standard Plus
Catalog No.	CBP10517
Lot No.	CBFA22102913
Format	Genomic DNA
DNA Change	c.1633G>A
AA Change	p.E545K
Mutation Type	Substitution - Missense
Zygotity	Homozygous
%AF	100.00%
Transcript	NM_006218.4
Chr position(GRCh37)	chr3:178936091
Copy Number	1
Ploidy	1
Buffer	Tris-EDTA
Storage Conditions	-20°C
Date of Manufacture	Oct.-29-2022
Expiry	36 months from the date of manufacture

Quality Control


Test Items	Standard	Results	Passed
Quantity	1.00 µg/vial*1 vial	1.00 µg/vial*1 vial	√
Concentration	40.00 ng/uL±10%	41.00 ng/uL by Qubit 3.0 detection	√
OD ₂₆₀ /OD ₂₈₀	1.7~2.1	1.90 by Nanodrop detection	√
DNA electrophoresis	No Degradation	 <p style="text-align: center;">M 1</p> <p>M: Marker 1: AI-Edigene® PIK3CA p.E545K Reference Standard Plus</p>	√
Genotype	Positive	 <p>By sanger sequencing</p>	√

%AF	100.00%	100.00% by DdPCR detection	√
-----	---------	----------------------------	---

Caution:

For research use only.

If you have any questions about the Certificate of Analysis, please contact us.

Certified by:  Date:****

 Date:****