



## CERTIFICATE OF ANALYSIS

### Product information

Name	CEBPA p.T337M Reference Standard
Catalog No.	CBP10546
Lot No.	CGD29108415
Format	Genomic DNA
DNA Change	c.1010C>T
AA Change	p.T337M
Mutation Type	Missense Mutation
Zygoty	Heterozygous
%AF	50.00%
Transcript	NM_004364.5
Chr position(GRCh37)	chr19:33792311
Copy Number	2
Ploidy	1.93
Buffer	Tris-EDTA
Storage Conditions	4°C
Date of Manufacture	Sep.-30-2021
Expiry	36 months from the date of manufacture


### Quality Control

Test Items	Standard	Results	Passed
Quantity	1.00 µg/vial*1 vial	1.00 µg/vial*1 vial	√
Concentration	40.00 ng/uL±10%	40.00 ng/uL by Qubit 3.0 detection	√
OD <sub>260</sub> /OD <sub>280</sub>	1.7~2.1	1.91 by Nanodrop detection	√
DNA electrophoresis	No Degradation	 <p>1kb ladder</p> <p>bp ng/5µl</p> <p>10000 40</p> <p>8000 40</p> <p>6000 40</p> <p>4000 40</p> <p>3000 40</p> <p>2000 100</p> <p>1500 40</p> <p>1000 40</p> <p>500 40</p> <p>1% agarose</p> <p>M 1</p> <p>M: Marker</p> <p>1: CEBPA p.T337M Reference Standard</p>	√
Genotype	Positive	 <p>110 120 130 140 150</p> <p>ACA GCT G A G C C G C G A A C T G G A C A C G C T G C G G G G C A T C T T C C G C C A G C</p> <p>ACA GCT G A G C C G C G A A C T G G A C A C G C T G C G G G G C A T C T T C C G C C A G C</p> <p>By sanger sequencing</p>	√

**Caution:**

For research use only.

If you have any questions about the Certificate of Analysis, please contact us.

Certified by:  Date:\*\*\*\*

 Date:\*\*\*\*