


## CERTIFICATE OF ANALYSIS

### Product information

Name	1p19q co-deletion Reference Standard
Catalog No.	CBP10557
Lot No.	GBD71907229
Format	Genomic DNA
Buffer	Tris-EDTA
Storage Conditions	4°C
Date of Manufacture	Jun.-30-2022
Expiry	36 months from the date of manufacture

### Quality Control

Test Items	Standard	Results	Passed
Quantity	1.00 µg/vial*1 vial	1.00 µg/vial*1 vial	√
Concentration	40 ng/uL±10%	43 ng/uL	√
OD <sub>260</sub> /OD <sub>280</sub>	1.7~2.1	1.86	√
DNA electrophoresis	No Degradation	 <p>           1kb ladder            bp ng/µl            10000 40            8000 40            6000 40            4000 40            3000 40            2000 100            1500 40            1000 40            500 40            1% agarose            M            M: Marker            1: 1p19q co-deletion Reference Standard         </p>	√

### Caution:

For research use only.

If you have any questions about the Certificate of Analysis, please contact us.

Certified by:  Date:\*\*\*\*

 Date:\*\*\*\*