Tel: 4008-750-250

QQ: 4008-750-250 Email: sales@cobioer.com



CERTIFICATE OF ANALYSIS

Product information

Name	TERT C250T Reference Standard_Skin			
Catalog No.	CBP10570			
Lot No.	CBFA22111685			
Format	Genomic DNA			
DNA Change	C250T			
AA Change	N/A			
Mutation Type	N/A			
Zygosity	Heterozygous			
%AF	66.67%			
Copy Number	3			
Ploidy	2.69			
Buffer	Tris-EDTA			
Storage Conditions	4°C			
Date of Manufacture	Nov16-2022			
Expiry	36 months from the date of manufacture			

Quality Control

Test Items	Standard	Results	Passed
Quantity	0.50 μg/vial*1 vial	0.50 μg/vial*1 vial	V
Concentration	40.00 ng/uL±10%	39.70 ng/uL by Qubit 3.0 detection	V
OD_{260}/OD_{280}	1.7~2.1	1.79 by Nanodrop detection	
DNA electrophoresis	No Degradation	1kb ladder	√
Genotype	Positive	240 250 260 270 220 66ACCCCCCCCCCCCCCCCCCCCCCCCCCCCCCCCCC	√
%AF	66.67%	69.32% By DdPCR Detection	√

Caution:

For research use only.

If you have any questions about the Certificate of Analysis, please contact us.





Certified by: Date:***

Date:***

Tel: 4008-750-250 QQ: 4008-750-250

Email: sales@cobioer.com