



## CERTIFICATE OF ANALYSIS

### Product information

Name	MSH6 Splice Site Reference Standard
Catalog No.	CBP10580
Lot No.	CBFA22120303
Format	Genomic DNA
DNA Change	c.4001+12_4001+15del
AA Change	N/A
Mutation Type	Splice Site
Zygoty	Homozygous
%AF	100%
Transcript	NM_000179.3
Chr position(GRCh37)	chr2-48033792-TAAC-
Copy Number	3
Ploidy	2.86
Buffer	Tris-EDTA
Storage Conditions	4°C
Date of Manufacture	Dec.-03-2022
Expiry	36 months from the date of manufacture


### Quality Control

Test Items	Standard	Results	Passed
Quantity	1.00 µg/vial*1 vial	1.00 µg/vial*1 vial	√
Concentration	40.00 ng/uL±10%	38.30 ng/uL by Qubit 3.0 detection	√
OD <sub>260</sub> /OD <sub>280</sub>	1.7~2.1	1.75 by Nanodrop detection	√
DNA electrophoresis	No Degradation	 <p>1kb ladder</p> <p>bp ng/5µl</p> <p>10000 40</p> <p>8000 40</p> <p>6000 40</p> <p>4000 40</p> <p>3000 40</p> <p>2000 100</p> <p>1500 40</p> <p>1000 40</p> <p>500 40</p> <p>1% agarose</p> <p>M 1</p> <p>M: Marker</p> <p>1: MSH6 Splice Site Reference Standard</p>	√
Genotype	Positive	 <p>410 420 430 440 450 460 470</p> <p>CAGTCCTACGATTATTCGGTAACTAACTA--TAAATGGAAATTATAACTAACTGACCTTAAAGT</p> <p>CAGTCCTACGATTATTCGGTAACTAACTA--TAAATGGAAATTATAACTAACTGACCTTAAAGT</p> <p>CAGTCCTACGATTATTCGGTAACTAACTA--TAAATGGAAATTATAACTAACTGACCTTAAAGT</p> <p>By sanger sequencing</p>	√

**Caution:**

For research use only.

If you have any questions about the Certificate of Analysis, please contact us.

Certified by:  Date:\*\*\*\*

 Date:\*\*\*\*