



CERTIFICATE OF ANALYSIS

Product information

Name	ESR1 p.Y537S Reference Standard
Catalog No.	CBP10603
Lot No.	CBFA23063001
Format	Genomic DNA
DNA Change	c.1610A>C
AA Change	p.Y537S
Mutation Type	Substitution - Missense
Zygosity	Homozygous
%AF	100.00%
Transcript	NM_001122740.2
Copy Number	3
Ploidy	2.97
Buffer	Tris-EDTA
Storage Conditions	4°C
Date of Manufacture	2023/06/30
Expiry	36 months from the date of manufacture

Quality Control

Test Items	Standard	Results	Passed
Quantity	1.00 µg/vial*20 vials	1.00 µg/vial*20 vials	√
Concentration	40.00 ng/uL±10%	40.30 ng/uL by Qubit 3.0 detection	√
OD ₂₆₀ /OD ₂₈₀	1.7~2.1	1.86 by Nanodrop detection	√
DNA electrophoresis	No Degradation	 <p>1kb ladder</p> <p>M: Marker 1: ESR1 p.Y537S Reference Standard</p>	√
Genotype	Positive	 <p>By sanger sequencing</p>	√

Caution:

For research use only.

If you have any questions about the Certificate of Analysis, please contact us.

Certified by:  Date:****

 Date:****