

# CERTIFICATE OF ANALYSIS

## General information

Name	Negative control for RNA
Catalog No.	CBP20000R
Lot No.	CBFA22122709
Format	RNA
Buffer	RNase-free H <sub>2</sub> O
Storage Conditions	-80°C
Date of Manufacture	2022/12/27
Expiry	12 months from the date of manufacture

## Quality Control

Test Items	Standard	Results	Passed
Quantity	1.00 µg/vial*1 vial	1.00 µg/vial*1 vial	√
Concentration	60.00 ng/uL±10%	60.00 ng/uL by Nanodrop Detection	√
OD <sub>260</sub> /OD <sub>280</sub>	1.8~2.2	2.06 by Nanodrop detection	√
RQN value	>9	9.92 by Qsep1 detection	√

### Caution:

For research use only.

If you have any questions about the Certificate of Analysis, please contact us.

Certified by:  Date:\*\*\*\*

 Date:\*\*\*\*