

CERTIFICATE OF ANALYSIS

General information

Name	Negative control for RNA		
Catalog No.	CBP20000R		
Lot No.	CBFA23061601		
Format	RNA		
Buffer	RNase-free H ₂ O		
Storage Conditions	-80°C		
Date of Manufacture 2023/06/16			
Expiry	12 months from the date of manufacture		

Ouality Control

Test Items	Standard	Results	Passed
Quantity	1.00 μg/vial*3 vials	1.00 μg/vial*3 vials	
Concentration	60.00 ng/uL±10%	60.00 ng/uL by Nanodrop Detection	
OD ₂₆₀ /OD ₂₈₀	1.8~2.2	2.06 by Nanodrop detection	
RQN value	>9	9.28 by Qsep1 detection	

Caution:

For research use only.

If you have any questions about the Certificate of Analysis, please contact us.

Certified by: Date:**** 1尊唱和 Date:****