


CERTIFICATE OF ANALYSIS

Product information

Name	CCDC6(E1)-RET(E12) Fusion
Catalog No.	CBP20014R
Lot No.	CBFA22122708
Format	RNA
Left Gene	CCDC6
Left Breakpoint	chr10:61665880:-
Right Gene	RET
Right Breakpoint	chr10:43612032:+
Ploidy	3.37
Buffer	RNase-free H ₂ O
Storage Conditions	-80°C
Date of Manufacture	Dec.-27-2022
Expiry	12 months from the date of manufacture

Quality Control

Test Items	Standard	Results	Passed
Quantity	1.00 µg/vial*1 vial	1.00 µg/vial*1 vial	√
Concentration	60.00 ng/uL±10%	63.10 ng/uL by Nanodrop detection	√
OD ₂₆₀ /OD ₂₈₀	1.8~2.2	1.94 by Nanodrop detection	√
RQN value	>9	9.1 by Qsep1 detection	√
Genotype	Positive	 By Sanger Sequencing	√
Copy number	Positive	1507.11 copies/ng by DdPCR detection	√

Caution:

For research use only.

If you have any questions about the Certificate of Analysis, please contact us.

Certified by:  Date:****

 Date:****