



CERTIFICATE OF ANALYSIS

CCDC6(E1)-RET(E12) Fusion

Product information

Name	CCDC6(E1)-RET(E12) Fusion
Catalog No.	CBP20014R
Lot No.	CGD2022061507
Format	RNA
Left Gene	CCDC6
Left Breakpoint	chr10:61665880:-
Right Gene	RET
Right Breakpoint	chr10:43612032:+
Ploidy	3.37
Buffer	RNase-free H2O
Storage Conditions	-80°C
Date of Manufacture	Jun.-15-2022
Expiry	12 months from the date of manufacture


Quality Control

Test Items	Standard	Results	Passed
Quantity	1ug/vial*1vial	1ug/vial*1vial	√
Concentration	Qubit 3.0	60 ng/uL±10%	√
OD ₂₆₀ /OD ₂₈₀	1.8~2.1	1.91	√
DNA electrophoresis	No Degradation	 <p>M 1 M: Marker 1: CCDC6(E1)-RET(E12) Fusion</p>	√
Genotype	Sanger Sequencing		√
Copy Number	DdPCR	1507.11 copies/ng	√

Caution:

For research use only.

If you have any questions about the Certificate of Analysis, please contact us.

Certified by:  Date:****

 Date:****