


CERTIFICATE OF ANALYSIS

Product information

Name	GOPC(E7)-ROS1(E35) Fusion
Catalog No.	CBP20021R
Lot No.	CGFA23032001
Format	RNA
Left Gene	GOPC(E7)
Left Breakpoint	chr6:117888017:-
Right Gene	ROS1(E35)
Right Breakpoint	chr6:117642557:-
Ploidy	2.18
Buffer	RNase-free H ₂ O
Storage Conditions	-80°C
Date of Manufacture	Mar.-20-2023
Expiry	12 months from the date of manufacture


Quality Control

Test Items	Standard	Results	Passed
Quantity	1.00 µg/vial*2 vials	1.00 µg/vial*2 vials	√
Concentration	60.00 ng/uL±10%	63.20 ng/uL by Nanodrop detection	√
OD ₂₆₀ /OD ₂₈₀	1.8~2.2	2.07 by Nanodrop detection	√
RQN value	>9	9.12 by Qsep1 detection	√
Genotype	Positive	 By Sanger Sequencing	√

Caution:

For research use only.

If you have any questions about the Certificate of Analysis, please contact us.

Certified by:  Date:****

 Date:****