Tel: 4008-750-250 QQ: 4008-750-250

Email: sales@cobioer.com



CERTIFICATE OF ANALYSIS

Product information

Name	SLC34A2(E4)-ROS1(E32) Translocation
Catalog No.	CBP20023D
Lot No.	CBFA22102604
Format	Genomic DNA
Left Gene	SLC34A2
Left Breakpoint	chr4:25666627:+
Left Gene Copy Number	4
Right Gene	ROS1
Right Breakpoint	chr6:117658307:-
Right Gene Copy Number	4
Ploidy	3.94
%AF	50.00%
Buffer	Tris-EDTA
Storage Conditions	-20°C
Date of Manufacture	Oct26-2022
Expiry	36 months from the date of manufacture

Quality Control

Quanty Control			
Test Items	Standard	Results	Passed
Quantity	1.00 µg/vial*2 vials	1.00 μg/vial*2 vials	
Concentration	40.00 ng/uL±10%	38.40 ng/uL by Qubit 3.0 detection	
OD_{260}/OD_{280}	1.7~2.1	1.86 by Nanodrop detection	
DNA electrophoresis	No Degradation	M 1 M: Marker 1: SLC34A2(E4)-ROS1(E32) Translocation	√
Genotype	Positive	By Sanger Sequencing	V
%AF	50.00%	49.83% by DdPCR detection	√





Caution:

For research use only.

If you have any questions about the Certificate of Analysis, please contact us.

Certified by: Date:***

1919 Date:***

Tel: 4008-750-250

QQ: 4008-750-250

Email: sales@cobioer.com